



ANALYTICAL REPORT

PREPARED FOR

Attn: Justin Huntley
Union County Water
500 N Main St.
Monroe, North Carolina 28112

Generated 9/13/2024 4:49:28 AM

JOB DESCRIPTION

PFAS - 533

JOB NUMBER

810-119317-1

Eurofins Eaton Analytical South Bend

Job Notes

This report may not be reproduced except in full, and with written approval from the laboratory. The results relate only to the samples tested. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

The test results in this report relate only to the samples as received by the laboratory and will meet all requirements of the methodology, with any exceptions noted. This report shall not be reproduced except in full, without the express written approval of the laboratory. All questions should be directed to the Eurofins Eaton Analytical, LLC Project Manager.

Authorization



Generated
9/13/2024 4:49:28 AM

Authorized for release by
Joe Mattheis, Project Manager I
Joe.Mattheis@et.eurofinsus.com
(574)233-4777



Table of Contents

| | |
|------------------------------------|----|
| Cover Page | 1 |
| Table of Contents | 3 |
| Definitions/Glossary | 4 |
| Case Narrative | 5 |
| Detection Summary | 6 |
| Client Sample Results | 7 |
| Isotope Dilution Summary | 11 |
| QC Sample Results | 12 |
| QC Association Summary | 16 |
| Lab Chronicle | 17 |
| Certification Summary | 18 |
| Method Summary | 19 |
| Sample Summary | 20 |
| Chain of Custody | 21 |
| Receipt Checklists | 22 |

Definitions/Glossary

Client: Union County Water
Project/Site: PFAS - 533

Job ID: 810-119317-1

Qualifiers

LCMS

| Qualifier | Qualifier Description |
|-----------|--|
| J | Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value. |

Glossary

| Abbreviation | These commonly used abbreviations may or may not be present in this report. |
|----------------|---|
| α | Listed under the "D" column to designate that the result is reported on a dry weight basis |
| %R | Percent Recovery |
| CFL | Contains Free Liquid |
| CFU | Colony Forming Unit |
| CNF | Contains No Free Liquid |
| DER | Duplicate Error Ratio (normalized absolute difference) |
| Dil Fac | Dilution Factor |
| DL | Detection Limit (DoD/DOE) |
| DL, RA, RE, IN | Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample |
| DLC | Decision Level Concentration (Radiochemistry) |
| EDL | Estimated Detection Limit (Dioxin) |
| LOD | Limit of Detection (DoD/DOE) |
| LOQ | Limit of Quantitation (DoD/DOE) |
| MCL | EPA recommended "Maximum Contaminant Level" |
| MDA | Minimum Detectable Activity (Radiochemistry) |
| MDC | Minimum Detectable Concentration (Radiochemistry) |
| MDL | Method Detection Limit |
| ML | Minimum Level (Dioxin) |
| MPN | Most Probable Number |
| MQL | Method Quantitation Limit |
| NC | Not Calculated |
| ND | Not Detected at the reporting limit (or MDL or EDL if shown) |
| NEG | Negative / Absent |
| POS | Positive / Present |
| PQL | Practical Quantitation Limit |
| PRES | Presumptive |
| QC | Quality Control |
| RER | Relative Error Ratio (Radiochemistry) |
| RL | Reporting Limit or Requested Limit (Radiochemistry) |
| RPD | Relative Percent Difference, a measure of the relative difference between two points |
| TEF | Toxicity Equivalent Factor (Dioxin) |
| TEQ | Toxicity Equivalent Quotient (Dioxin) |
| TNTC | Too Numerous To Count |

Case Narrative

Client: Union County Water
Project: PFAS - 533

Job ID: 810-119317-1

Job ID: 810-119317-1

Eurofins Eaton Analytical South Bend

Job Narrative 810-119317-1

Analytical test results meet all requirements of the associated regulatory program listed on the Accreditation/Certification Summary Page unless otherwise noted under the individual analysis. Data qualifiers and/or narrative comments are included to explain any exceptions, if applicable.

- Matrix QC may not be reported if insufficient sample is provided or site-specific QC samples were not submitted. In these situations, to demonstrate precision and accuracy at a batch level, a LCS/LCSD may be performed, unless otherwise specified in the method.
- Surrogate and/or isotope dilution analyte recoveries (if applicable) which are outside of the QC window are confirmed unless attributed to a dilution or otherwise noted in the narrative.

Regulated compliance samples (e.g. SDWA, NPDES) must comply with the associated agency requirements/permits.

Receipt

The samples were received on 9/10/2024 9:30 AM. Unless otherwise noted below, the samples arrived in good condition, and, where required, properly preserved and on ice. The temperature of the cooler at receipt time was 1.4°C.

PFAS

Method 533: The pH of the following sample was adjusted to pH 7.5 in the laboratory: J18 - Rehoboth ARV (810-119317-1)

Method 533: The pH of the following sample was adjusted to pH 7.5 in the laboratory: Y01 - Yadkin Finished Water (810-119317-2)

Method 533: The pH of the following sample was adjusted to pH 7.5 in the laboratory: Y02 - Yadkin Raw Water (810-119317-3)

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

Detection Summary

Client: Union County Water
Project/Site: PFAS - 533

Job ID: 810-119317-1

Client Sample ID: J18 - Rehoboth ARV

Lab Sample ID: 810-119317-1

| Analyte | Result | Qualifier | RL | MDL | Unit | Dil Fac | D | Method | Prep Type |
|-------------------------------------|--------|-----------|-----|-----|------|---------|---|--------|-----------|
| Perfluorobutanoic acid (PFBA) | 4.0 | | 1.9 | | ng/L | 1 | | 533 | Total/NA |
| Perfluoropentanoic acid (PFPeA) | 7.9 | | 1.9 | | ng/L | 1 | | 533 | Total/NA |
| Perfluorohexanoic acid (PFHxA) | 6.9 | | 1.9 | | ng/L | 1 | | 533 | Total/NA |
| Perfluoroheptanoic acid (PFHpA) | 2.2 | | 1.9 | | ng/L | 1 | | 533 | Total/NA |
| Perfluorooctanoic acid (PFOA) | 3.8 | | 1.9 | | ng/L | 1 | | 533 | Total/NA |
| Perfluorobutanesulfonic acid (PFBS) | 2.2 | | 1.9 | | ng/L | 1 | | 533 | Total/NA |
| Perfluorooctanesulfonic acid (PFOS) | 3.2 | | 1.9 | | ng/L | 1 | | 533 | Total/NA |

Client Sample ID: Y01 - Yadkin Finished Water

Lab Sample ID: 810-119317-2

| Analyte | Result | Qualifier | RL | MDL | Unit | Dil Fac | D | Method | Prep Type |
|---------------------------------|--------|-----------|-----|-----|------|---------|---|--------|-----------|
| Perfluorobutanoic acid (PFBA) | 2.7 | | 1.9 | | ng/L | 1 | | 533 | Total/NA |
| Perfluoropentanoic acid (PFPeA) | 2.6 | | 1.9 | | ng/L | 1 | | 533 | Total/NA |
| Perfluorohexanoic acid (PFHxA) | 2.3 | | 1.9 | | ng/L | 1 | | 533 | Total/NA |

Client Sample ID: Y02 - Yadkin Raw Water

Lab Sample ID: 810-119317-3

| Analyte | Result | Qualifier | RL | MDL | Unit | Dil Fac | D | Method | Prep Type |
|-------------------------------------|--------|-----------|-----|-----|------|---------|---|--------|-----------|
| Perfluorobutanoic acid (PFBA) | 3.0 | | 1.9 | | ng/L | 1 | | 533 | Total/NA |
| Perfluoropentanoic acid (PFPeA) | 3.4 | | 1.9 | | ng/L | 1 | | 533 | Total/NA |
| Perfluorohexanoic acid (PFHxA) | 3.7 | | 1.9 | | ng/L | 1 | | 533 | Total/NA |
| Perfluorooctanoic acid (PFOA) | 2.9 | | 1.9 | | ng/L | 1 | | 533 | Total/NA |
| Perfluorobutanesulfonic acid (PFBS) | 2.0 | | 1.9 | | ng/L | 1 | | 533 | Total/NA |
| Perfluorooctanesulfonic acid (PFOS) | 5.0 | | 1.9 | | ng/L | 1 | | 533 | Total/NA |

This Detection Summary does not include radiochemical test results.

Eurofins Eaton Analytical South Bend

Client Sample Results

Client: Union County Water
Project/Site: PFAS - 533

Job ID: 810-119317-1

Client Sample ID: J18 - Rehoboth ARV

Lab Sample ID: 810-119317-1

Date Collected: 09/09/24 11:10

Matrix: Drinking Water

Date Received: 09/10/24 09:30

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

| Analyte | Result | Qualifier | RL | MDL | Unit | D | Prepared | Analyzed | Dil Fac |
|--|--------|-----------|-----|-----|------|---|----------------|----------------|---------|
| Perfluorobutanoic acid (PFBA) | 4.0 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 03:01 | 1 |
| Perfluoropentanoic acid (PFPeA) | 7.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 03:01 | 1 |
| Perfluorohexanoic acid (PFHxA) | 6.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 03:01 | 1 |
| Perfluoroheptanoic acid (PFHpA) | 2.2 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 03:01 | 1 |
| Perfluorooctanoic acid (PFOA) | 3.8 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 03:01 | 1 |
| Perfluorononanoic acid (PFNA) | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 03:01 | 1 |
| Perfluorodecanoic acid (PFDA) | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 03:01 | 1 |
| Perfluoroundecanoic acid (PFUnA) | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 03:01 | 1 |
| Perfluorododecanoic acid (PFDoA) | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 03:01 | 1 |
| Perfluorobutanesulfonic acid (PFBS) | 2.2 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 03:01 | 1 |
| Perfluoropentanesulfonic acid (PFPeS) | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 03:01 | 1 |
| Perfluorohexanesulfonic acid (PFHxS) | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 03:01 | 1 |
| Perfluoroheptanesulfonic acid (PFHpS) | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 03:01 | 1 |
| Perfluorooctanesulfonic acid (PFOS) | 3.2 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 03:01 | 1 |
| Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA) | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 03:01 | 1 |
| 1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS) | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 03:01 | 1 |
| 1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS) | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 03:01 | 1 |
| 1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS) | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 03:01 | 1 |
| Hexafluoropropylene Oxide Dimer Acid (HFPO-DA) | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 03:01 | 1 |
| 4,8-Dioxa-3H-perfluorononanoic acid (ADONA) | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 03:01 | 1 |
| 9-Chlorohexadecafluoro-3-oxanonan e-1-sulfonic acid | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 03:01 | 1 |
| 11-Chloroeicosafluoro-3-oxaundecan e-1-sulfonic acid | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 03:01 | 1 |
| Perfluoro(4-methoxybutanoic acid) | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 03:01 | 1 |
| Perfluoro-3-methoxypropanoic acid (PFMPA) | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 03:01 | 1 |
| Perfluoro-3,6-dioxaheptanoic acid | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 03:01 | 1 |

| Isotope Dilution | %Recovery | Qualifier | Limits | Prepared | Analyzed | Dil Fac |
|------------------|-----------|-----------|----------|----------------|----------------|---------|
| 13C4 PFBA | 96 | | 50 - 200 | 09/11/24 07:53 | 09/12/24 03:01 | 1 |
| 13C5 PFPeA | 115 | | 50 - 200 | 09/11/24 07:53 | 09/12/24 03:01 | 1 |
| 13C5 PFHxA | 95 | | 50 - 200 | 09/11/24 07:53 | 09/12/24 03:01 | 1 |
| 13C4 PFHpA | 92 | | 50 - 200 | 09/11/24 07:53 | 09/12/24 03:01 | 1 |
| 13C8 PFOA | 82 | | 50 - 200 | 09/11/24 07:53 | 09/12/24 03:01 | 1 |
| 13C9 PFNA | 75 | | 50 - 200 | 09/11/24 07:53 | 09/12/24 03:01 | 1 |
| 13C6 PFDA | 72 | | 50 - 200 | 09/11/24 07:53 | 09/12/24 03:01 | 1 |
| 13C7 PFUnA | 73 | | 50 - 200 | 09/11/24 07:53 | 09/12/24 03:01 | 1 |
| 13C2 PFDoA | 73 | | 50 - 200 | 09/11/24 07:53 | 09/12/24 03:01 | 1 |
| 13C3 HFPO-DA | 93 | | 50 - 200 | 09/11/24 07:53 | 09/12/24 03:01 | 1 |
| 13C3 PFBS | 102 | | 50 - 200 | 09/11/24 07:53 | 09/12/24 03:01 | 1 |
| 13C8 PFOS | 95 | | 50 - 200 | 09/11/24 07:53 | 09/12/24 03:01 | 1 |
| 13C2-4:2-FTS | 115 | | 50 - 200 | 09/11/24 07:53 | 09/12/24 03:01 | 1 |

Eurofins Eaton Analytical South Bend

Client Sample Results

Client: Union County Water
Project/Site: PFAS - 533

Job ID: 810-119317-1

Client Sample ID: J18 - Rehoboth ARV

Lab Sample ID: 810-119317-1

Date Collected: 09/09/24 11:10

Matrix: Drinking Water

Date Received: 09/10/24 09:30

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

| Isotope Dilution | %Recovery | Qualifier | Limits | Prepared | Analyzed | Dil Fac |
|------------------|-----------|-----------|----------|----------------|----------------|---------|
| 13C2-6:2-FTS | 104 | | 50 - 200 | 09/11/24 07:53 | 09/12/24 03:01 | 1 |
| 13C2-8:2-FTS | 101 | | 50 - 200 | 09/11/24 07:53 | 09/12/24 03:01 | 1 |
| 13C3 PFHxS | 99 | | 50 - 200 | 09/11/24 07:53 | 09/12/24 03:01 | 1 |

Client Sample ID: Y01 - Yadkin Finished Water

Lab Sample ID: 810-119317-2

Date Collected: 09/09/24 11:57

Matrix: Drinking Water

Date Received: 09/10/24 09:30

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

| Analyte | Result | Qualifier | RL | MDL | Unit | D | Prepared | Analyzed | Dil Fac |
|--|--------|-----------|-----|-----|------|---|----------------|----------------|---------|
| Perfluorobutanoic acid (PFBA) | 2.7 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 03:15 | 1 |
| Perfluoropentanoic acid (PFPeA) | 2.6 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 03:15 | 1 |
| Perfluorohexanoic acid (PFHxA) | 2.3 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 03:15 | 1 |
| Perfluoroheptanoic acid (PFHpA) | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 03:15 | 1 |
| Perfluorooctanoic acid (PFOA) | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 03:15 | 1 |
| Perfluorononanoic acid (PFNA) | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 03:15 | 1 |
| Perfluorodecanoic acid (PFDA) | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 03:15 | 1 |
| Perfluoroundecanoic acid (PFUnA) | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 03:15 | 1 |
| Perfluorododecanoic acid (PFDoA) | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 03:15 | 1 |
| Perfluorobutanesulfonic acid (PFBS) | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 03:15 | 1 |
| Perfluoropentanesulfonic acid (PFPeS) | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 03:15 | 1 |
| Perfluorohexanesulfonic acid (PFHxS) | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 03:15 | 1 |
| Perfluoroheptanesulfonic acid (PFHpS) | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 03:15 | 1 |
| Perfluorooctanesulfonic acid (PFOS) | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 03:15 | 1 |
| Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA) | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 03:15 | 1 |
| 1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS) | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 03:15 | 1 |
| 1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS) | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 03:15 | 1 |
| 1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS) | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 03:15 | 1 |
| Hexafluoropropylene Oxide Dimer Acid (HFPO-DA) | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 03:15 | 1 |
| 4,8-Dioxa-3H-perfluorononanoic acid (ADONA) | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 03:15 | 1 |
| 9-Chlorohexadecafluoro-3-oxanonan e-1-sulfonic acid | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 03:15 | 1 |
| 11-Chloroeicosafluoro-3-oxaundecan e-1-sulfonic acid | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 03:15 | 1 |
| Perfluoro(4-methoxybutanoic acid) | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 03:15 | 1 |
| Perfluoro-3-methoxypropanoic acid (PFMPA) | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 03:15 | 1 |
| Perfluoro-3,6-dioxaheptanoic acid | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 03:15 | 1 |

| Isotope Dilution | %Recovery | Qualifier | Limits | Prepared | Analyzed | Dil Fac |
|------------------|-----------|-----------|----------|----------------|----------------|---------|
| 13C4 PFBA | 99 | | 50 - 200 | 09/11/24 07:53 | 09/12/24 03:15 | 1 |
| 13C5 PFPeA | 108 | | 50 - 200 | 09/11/24 07:53 | 09/12/24 03:15 | 1 |
| 13C5 PFHxA | 95 | | 50 - 200 | 09/11/24 07:53 | 09/12/24 03:15 | 1 |
| 13C4 PFHpA | 92 | | 50 - 200 | 09/11/24 07:53 | 09/12/24 03:15 | 1 |
| 13C8 PFOA | 85 | | 50 - 200 | 09/11/24 07:53 | 09/12/24 03:15 | 1 |

Eurofins Eaton Analytical South Bend

Client Sample Results

Client: Union County Water
Project/Site: PFAS - 533

Job ID: 810-119317-1

Client Sample ID: Y01 - Yadkin Finished Water

Lab Sample ID: 810-119317-2

Date Collected: 09/09/24 11:57

Matrix: Drinking Water

Date Received: 09/10/24 09:30

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

| Isotope Dilution | %Recovery | Qualifier | Limits | Prepared | Analyzed | Dil Fac |
|------------------|-----------|-----------|----------|----------------|----------------|---------|
| 13C9 PFNA | 81 | | 50 - 200 | 09/11/24 07:53 | 09/12/24 03:15 | 1 |
| 13C6 PFDA | 79 | | 50 - 200 | 09/11/24 07:53 | 09/12/24 03:15 | 1 |
| 13C7 PFUnA | 81 | | 50 - 200 | 09/11/24 07:53 | 09/12/24 03:15 | 1 |
| 13C2 PFDoA | 80 | | 50 - 200 | 09/11/24 07:53 | 09/12/24 03:15 | 1 |
| 13C3 HFPO-DA | 93 | | 50 - 200 | 09/11/24 07:53 | 09/12/24 03:15 | 1 |
| 13C3 PFBS | 102 | | 50 - 200 | 09/11/24 07:53 | 09/12/24 03:15 | 1 |
| 13C8 PFOS | 101 | | 50 - 200 | 09/11/24 07:53 | 09/12/24 03:15 | 1 |
| 13C2-4:2-FTS | 103 | | 50 - 200 | 09/11/24 07:53 | 09/12/24 03:15 | 1 |
| 13C2-6:2-FTS | 103 | | 50 - 200 | 09/11/24 07:53 | 09/12/24 03:15 | 1 |
| 13C2-8:2-FTS | 100 | | 50 - 200 | 09/11/24 07:53 | 09/12/24 03:15 | 1 |
| 13C3 PFHxS | 100 | | 50 - 200 | 09/11/24 07:53 | 09/12/24 03:15 | 1 |

Client Sample ID: Y02 - Yadkin Raw Water

Lab Sample ID: 810-119317-3

Date Collected: 09/09/24 12:04

Matrix: Drinking Water

Date Received: 09/10/24 09:30

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

| Analyte | Result | Qualifier | RL | MDL | Unit | D | Prepared | Analyzed | Dil Fac |
|---|--------|-----------|-----|-----|------|---|----------------|----------------|---------|
| Perfluorobutanoic acid (PFBA) | 3.0 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 04:22 | 1 |
| Perfluoropentanoic acid (PFPeA) | 3.4 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 04:22 | 1 |
| Perfluorohexanoic acid (PFHxA) | 3.7 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 04:22 | 1 |
| Perfluoroheptanoic acid (PFHpA) | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 04:22 | 1 |
| Perfluorooctanoic acid (PFOA) | 2.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 04:22 | 1 |
| Perfluorononanoic acid (PFNA) | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 04:22 | 1 |
| Perfluorodecanoic acid (PFDA) | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 04:22 | 1 |
| Perfluoroundecanoic acid (PFUnA) | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 04:22 | 1 |
| Perfluorododecanoic acid (PFDoA) | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 04:22 | 1 |
| Perfluorobutanesulfonic acid (PFBS) | 2.0 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 04:22 | 1 |
| Perfluoropentanesulfonic acid (PFPeS) | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 04:22 | 1 |
| Perfluorohexanesulfonic acid (PFHxS) | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 04:22 | 1 |
| Perfluoroheptanesulfonic acid (PFHpS) | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 04:22 | 1 |
| Perfluorooctanesulfonic acid (PFOS) | 5.0 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 04:22 | 1 |
| Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA) | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 04:22 | 1 |
| 1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS) | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 04:22 | 1 |
| 1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS) | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 04:22 | 1 |
| 1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS) | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 04:22 | 1 |
| Hexafluoropropylene Oxide Dimer Acid (HFPO-DA) | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 04:22 | 1 |
| 4,8-Dioxa-3H-perfluorononanoic acid (ADONA) | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 04:22 | 1 |
| 9-Chlorohexadecafluoro-3-oxanonan e-1-sulfonic acid | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 04:22 | 1 |
| 11-Chloroeicosafuoro-3-oxaundecan e-1-sulfonic acid | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 04:22 | 1 |
| Perfluoro(4-methoxybutanoic acid) | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 04:22 | 1 |

Eurofins Eaton Analytical South Bend

Client Sample Results

Client: Union County Water
 Project/Site: PFAS - 533

Job ID: 810-119317-1

Client Sample ID: Y02 - Yadkin Raw Water

Lab Sample ID: 810-119317-3

Date Collected: 09/09/24 12:04

Matrix: Drinking Water

Date Received: 09/10/24 09:30

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

| Analyte | Result | Qualifier | RL | MDL | Unit | D | Prepared | Analyzed | Dil Fac |
|---|-----------|-----------|----------|-----|------|---|----------------|----------------|---------|
| Perfluoro-3-methoxypropanoic acid (PFMPA) | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 04:22 | 1 |
| Perfluoro-3,6-dioxaheptanoic acid | <1.9 | | 1.9 | | ng/L | | 09/11/24 07:53 | 09/12/24 04:22 | 1 |
| Isotope Dilution | %Recovery | Qualifier | Limits | | | | Prepared | Analyzed | Dil Fac |
| 13C4 PFBA | 98 | | 50 - 200 | | | | 09/11/24 07:53 | 09/12/24 04:22 | 1 |
| 13C5 PFPeA | 138 | | 50 - 200 | | | | 09/11/24 07:53 | 09/12/24 04:22 | 1 |
| 13C5 PFHxA | 97 | | 50 - 200 | | | | 09/11/24 07:53 | 09/12/24 04:22 | 1 |
| 13C4 PFHpA | 93 | | 50 - 200 | | | | 09/11/24 07:53 | 09/12/24 04:22 | 1 |
| 13C8 PFOA | 90 | | 50 - 200 | | | | 09/11/24 07:53 | 09/12/24 04:22 | 1 |
| 13C9 PFNA | 90 | | 50 - 200 | | | | 09/11/24 07:53 | 09/12/24 04:22 | 1 |
| 13C6 PFDA | 87 | | 50 - 200 | | | | 09/11/24 07:53 | 09/12/24 04:22 | 1 |
| 13C7 PFUnA | 87 | | 50 - 200 | | | | 09/11/24 07:53 | 09/12/24 04:22 | 1 |
| 13C2 PFDoA | 93 | | 50 - 200 | | | | 09/11/24 07:53 | 09/12/24 04:22 | 1 |
| 13C3 HFPO-DA | 94 | | 50 - 200 | | | | 09/11/24 07:53 | 09/12/24 04:22 | 1 |
| 13C3 PFBS | 103 | | 50 - 200 | | | | 09/11/24 07:53 | 09/12/24 04:22 | 1 |
| 13C8 PFOS | 97 | | 50 - 200 | | | | 09/11/24 07:53 | 09/12/24 04:22 | 1 |
| 13C2-4:2-FTS | 124 | | 50 - 200 | | | | 09/11/24 07:53 | 09/12/24 04:22 | 1 |
| 13C2-6:2-FTS | 102 | | 50 - 200 | | | | 09/11/24 07:53 | 09/12/24 04:22 | 1 |
| 13C2-8:2-FTS | 101 | | 50 - 200 | | | | 09/11/24 07:53 | 09/12/24 04:22 | 1 |
| 13C3 PFHxS | 98 | | 50 - 200 | | | | 09/11/24 07:53 | 09/12/24 04:22 | 1 |

Isotope Dilution Summary

Client: Union County Water
Project/Site: PFAS - 533

Job ID: 810-119317-1

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Matrix: Drinking Water

Prep Type: Total/NA

| | | Percent Isotope Dilution Recovery (Acceptance Limits) | | | | | | | |
|---------------------|-----------------------------|---|-------------------|---------------------|--------------------|--------------------|--------------------|--------------------|---------------------|
| Lab Sample ID | Client Sample ID | PFBA (50-200) | PFPeA (50-200) | 13C5PHA (50-200) | C4PFHA (50-200) | C8PFOA (50-200) | C9PFNA (50-200) | C6PFDA (50-200) | 13C7PUA (50-200) |
| 810-119317-1 | J18 - Rehoboth ARV | 96 | 115 | 95 | 92 | 82 | 75 | 72 | 73 |
| 810-119317-2 | Y01 - Yadkin Finished Water | 99 | 108 | 95 | 92 | 85 | 81 | 79 | 81 |
| 810-119317-3 | Y02 - Yadkin Raw Water | 98 | 138 | 97 | 93 | 90 | 90 | 87 | 87 |
| LCS 810-114219/3-A | Lab Control Sample | 83 | 83 | 80 | 82 | 87 | 85 | 83 | 82 |
| LLCS 810-114219/2-A | Lab Control Sample | 69 | 70 | 71 | 72 | 76 | 80 | 79 | 80 |
| MBL 810-114219/1-A | Method Blank | 62 | 63 | 63 | 65 | 67 | 70 | 71 | 73 |

| | | Percent Isotope Dilution Recovery (Acceptance Limits) | | | | | | | |
|---------------------|-----------------------------|---|--------------------|--------------------|--------------------|-------------------|-------------------|-------------------|--------------------|
| Lab Sample ID | Client Sample ID | PFD _o A (50-200) | HFPODA (50-200) | C3PFBS (50-200) | C8PFOS (50-200) | 42FTS (50-200) | 62FTS (50-200) | 82FTS (50-200) | C3PFHS (50-200) |
| 810-119317-1 | J18 - Rehoboth ARV | 73 | 93 | 102 | 95 | 115 | 104 | 101 | 99 |
| 810-119317-2 | Y01 - Yadkin Finished Water | 80 | 93 | 102 | 101 | 103 | 103 | 100 | 100 |
| 810-119317-3 | Y02 - Yadkin Raw Water | 93 | 94 | 103 | 97 | 124 | 102 | 101 | 98 |
| LCS 810-114219/3-A | Lab Control Sample | 80 | 78 | 97 | 100 | 104 | 111 | 105 | 97 |
| LLCS 810-114219/2-A | Lab Control Sample | 82 | 66 | 98 | 98 | 94 | 94 | 98 | 101 |
| MBL 810-114219/1-A | Method Blank | 76 | 59 | 101 | 99 | 91 | 94 | 99 | 103 |

Surrogate Legend

- PFBA = 13C4 PFBA
- PFPeA = 13C5 PFPeA
- 13C5PHA = 13C5 PFHxA
- C4PFHA = 13C4 PFHpA
- C8PFOA = 13C8 PFOA
- C9PFNA = 13C9 PFNA
- C6PFDA = 13C6 PFDA
- 13C7PUA = 13C7 PFUnA
- PFD_oA = 13C2 PFD_oA
- HFPODA = 13C3 HFPO-DA
- C3PFBS = 13C3 PFBS
- C8PFOS = 13C8 PFOS
- 42FTS = 13C2-4:2-FTS
- 62FTS = 13C2-6:2-FTS
- 82FTS = 13C2-8:2-FTS
- C3PFHS = 13C3 PFHxS

QC Sample Results

Client: Union County Water
Project/Site: PFAS - 533

Job ID: 810-119317-1

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Lab Sample ID: MBL 810-114219/1-A
Matrix: Drinking Water
Analysis Batch: 114282

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 114219

| Analyte | MBL Result | MBL Qualifier | RL | MDL | Unit | D | Prepared | Analyzed | Dil Fac |
|--|------------|---------------|-----|-----|------|---|----------------|----------------|---------|
| Perfluorobutanoic acid (PFBA) | <0.52 | | 2.0 | | ng/L | | 09/11/24 07:53 | 09/11/24 21:26 | 1 |
| Perfluoropentanoic acid (PFPeA) | <0.38 | | 2.0 | | ng/L | | 09/11/24 07:53 | 09/11/24 21:26 | 1 |
| Perfluorohexanoic acid (PFHxA) | <0.42 | | 2.0 | | ng/L | | 09/11/24 07:53 | 09/11/24 21:26 | 1 |
| Perfluoroheptanoic acid (PFHpA) | <0.40 | | 2.0 | | ng/L | | 09/11/24 07:53 | 09/11/24 21:26 | 1 |
| Perfluorooctanoic acid (PFOA) | <0.38 | | 2.0 | | ng/L | | 09/11/24 07:53 | 09/11/24 21:26 | 1 |
| Perfluorononanoic acid (PFNA) | <0.38 | | 2.0 | | ng/L | | 09/11/24 07:53 | 09/11/24 21:26 | 1 |
| Perfluorodecanoic acid (PFDA) | <0.36 | | 2.0 | | ng/L | | 09/11/24 07:53 | 09/11/24 21:26 | 1 |
| Perfluoroundecanoic acid (PFUnA) | <0.38 | | 2.0 | | ng/L | | 09/11/24 07:53 | 09/11/24 21:26 | 1 |
| Perfluorododecanoic acid (PFDoA) | <0.35 | | 2.0 | | ng/L | | 09/11/24 07:53 | 09/11/24 21:26 | 1 |
| Perfluorobutanesulfonic acid (PFBS) | <0.42 | | 2.0 | | ng/L | | 09/11/24 07:53 | 09/11/24 21:26 | 1 |
| Perfluoropentanesulfonic acid (PFPeS) | <0.37 | | 2.0 | | ng/L | | 09/11/24 07:53 | 09/11/24 21:26 | 1 |
| Perfluorohexanesulfonic acid (PFHxS) | <0.39 | | 2.0 | | ng/L | | 09/11/24 07:53 | 09/11/24 21:26 | 1 |
| Perfluoroheptanesulfonic acid (PFHpS) | <0.44 | | 2.0 | | ng/L | | 09/11/24 07:53 | 09/11/24 21:26 | 1 |
| Perfluorooctanesulfonic acid (PFOS) | <0.39 | | 2.0 | | ng/L | | 09/11/24 07:53 | 09/11/24 21:26 | 1 |
| Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA) | <0.45 | | 2.0 | | ng/L | | 09/11/24 07:53 | 09/11/24 21:26 | 1 |
| 1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS) | <0.56 | | 2.0 | | ng/L | | 09/11/24 07:53 | 09/11/24 21:26 | 1 |
| 1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS) | <0.68 | | 2.0 | | ng/L | | 09/11/24 07:53 | 09/11/24 21:26 | 1 |
| 1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS) | <0.57 | | 2.0 | | ng/L | | 09/11/24 07:53 | 09/11/24 21:26 | 1 |
| Hexafluoropropylene Oxide Dimer Acid (HFPO-DA) | <0.53 | | 2.0 | | ng/L | | 09/11/24 07:53 | 09/11/24 21:26 | 1 |
| 4,8-Dioxa-3H-perfluorononanoic acid (ADONA) | <0.40 | | 2.0 | | ng/L | | 09/11/24 07:53 | 09/11/24 21:26 | 1 |
| 9-Chlorohexadecafluoro-3-oxanonan e-1-sulfonic acid | <0.45 | | 2.0 | | ng/L | | 09/11/24 07:53 | 09/11/24 21:26 | 1 |
| 11-Chloroeicosafluoro-3-oxaundecan e-1-sulfonic acid | <0.51 | | 2.0 | | ng/L | | 09/11/24 07:53 | 09/11/24 21:26 | 1 |
| Perfluoro(4-methoxybutanoic acid) | <0.35 | | 2.0 | | ng/L | | 09/11/24 07:53 | 09/11/24 21:26 | 1 |
| Perfluoro-3-methoxypropanoic acid (PFMPA) | <0.32 | | 2.0 | | ng/L | | 09/11/24 07:53 | 09/11/24 21:26 | 1 |
| Perfluoro-3,6-dioxaheptanoic acid | <0.93 | | 2.0 | | ng/L | | 09/11/24 07:53 | 09/11/24 21:26 | 1 |

| Isotope Dilution | MBL %Recovery | MBL Qualifier | Limits | Prepared | Analyzed | Dil Fac |
|------------------|---------------|---------------|----------|----------------|----------------|---------|
| 13C4 PFBA | 62 | | 50 - 200 | 09/11/24 07:53 | 09/11/24 21:26 | 1 |
| 13C5 PFPeA | 63 | | 50 - 200 | 09/11/24 07:53 | 09/11/24 21:26 | 1 |
| 13C5 PFHxA | 63 | | 50 - 200 | 09/11/24 07:53 | 09/11/24 21:26 | 1 |
| 13C4 PFHpA | 65 | | 50 - 200 | 09/11/24 07:53 | 09/11/24 21:26 | 1 |
| 13C8 PFOA | 67 | | 50 - 200 | 09/11/24 07:53 | 09/11/24 21:26 | 1 |
| 13C9 PFNA | 70 | | 50 - 200 | 09/11/24 07:53 | 09/11/24 21:26 | 1 |
| 13C6 PFDA | 71 | | 50 - 200 | 09/11/24 07:53 | 09/11/24 21:26 | 1 |
| 13C7 PFUnA | 73 | | 50 - 200 | 09/11/24 07:53 | 09/11/24 21:26 | 1 |
| 13C2 PFDoA | 76 | | 50 - 200 | 09/11/24 07:53 | 09/11/24 21:26 | 1 |
| 13C3 HFPO-DA | 59 | | 50 - 200 | 09/11/24 07:53 | 09/11/24 21:26 | 1 |
| 13C3 PFBS | 101 | | 50 - 200 | 09/11/24 07:53 | 09/11/24 21:26 | 1 |
| 13C8 PFOS | 99 | | 50 - 200 | 09/11/24 07:53 | 09/11/24 21:26 | 1 |
| 13C2-4:2-FTS | 91 | | 50 - 200 | 09/11/24 07:53 | 09/11/24 21:26 | 1 |

Eurofins Eaton Analytical South Bend

QC Sample Results

Client: Union County Water
Project/Site: PFAS - 533

Job ID: 810-119317-1

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: MBL 810-114219/1-A
Matrix: Drinking Water
Analysis Batch: 114282

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 114219

| Isotope Dilution | MBL MBL | | Limits | Prepared | Analyzed | Dil Fac |
|------------------|-----------|-----------|----------|----------------|----------------|---------|
| | %Recovery | Qualifier | | | | |
| 13C2-6:2-FTS | 94 | | 50 - 200 | 09/11/24 07:53 | 09/11/24 21:26 | 1 |
| 13C2-8:2-FTS | 99 | | 50 - 200 | 09/11/24 07:53 | 09/11/24 21:26 | 1 |
| 13C3 PFHxS | 103 | | 50 - 200 | 09/11/24 07:53 | 09/11/24 21:26 | 1 |

Lab Sample ID: LCS 810-114219/3-A
Matrix: Drinking Water
Analysis Batch: 114282

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 114219

| Analyte | Spike Added | LCS Result | LCS Qualifier | Unit | D | %Rec | %Rec Limits |
|---|-------------|------------|---------------|------|---|------|-------------|
| | | | | | | | |
| Perfluoropentanoic acid (PFPeA) | 200 | 194 | | ng/L | | 97 | 70 - 130 |
| Perfluorohexanoic acid (PFHxA) | 200 | 198 | | ng/L | | 99 | 70 - 130 |
| Perfluoroheptanoic acid (PFHpA) | 200 | 197 | | ng/L | | 99 | 70 - 130 |
| Perfluorooctanoic acid (PFOA) | 200 | 192 | | ng/L | | 96 | 70 - 130 |
| Perfluorononanoic acid (PFNA) | 200 | 192 | | ng/L | | 96 | 70 - 130 |
| Perfluorodecanoic acid (PFDA) | 200 | 196 | | ng/L | | 98 | 70 - 130 |
| Perfluoroundecanoic acid (PFUnA) | 200 | 199 | | ng/L | | 100 | 70 - 130 |
| Perfluorododecanoic acid (PFDoA) | 200 | 199 | | ng/L | | 100 | 70 - 130 |
| Perfluorobutanesulfonic acid (PFBS) | 178 | 176 | | ng/L | | 99 | 70 - 130 |
| Perfluoropentanesulfonic acid (PFPeS) | 188 | 180 | | ng/L | | 96 | 70 - 130 |
| Perfluorohexanesulfonic acid (PFHxS) | 183 | 178 | | ng/L | | 97 | 70 - 130 |
| Perfluoroheptanesulfonic acid (PFHpS) | 191 | 189 | | ng/L | | 99 | 70 - 130 |
| Perfluorooctanesulfonic acid (PFOS) | 186 | 179 | | ng/L | | 96 | 70 - 130 |
| Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA) | 178 | 171 | | ng/L | | 96 | 70 - 130 |
| 1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS) | 188 | 190 | | ng/L | | 102 | 70 - 130 |
| 1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS) | 190 | 195 | | ng/L | | 102 | 70 - 130 |
| 1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS) | 192 | 195 | | ng/L | | 102 | 70 - 130 |
| Hexafluoropropylene Oxide Dimer Acid (HFPO-DA) | 200 | 200 | | ng/L | | 100 | 70 - 130 |
| 4,8-Dioxa-3H-perfluorononanoic acid (ADONA) | 189 | 178 | | ng/L | | 94 | 70 - 130 |
| 9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid | 187 | 203 | | ng/L | | 109 | 70 - 130 |
| 11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid | 189 | 181 | | ng/L | | 96 | 70 - 130 |
| Perfluoro(4-methoxybutanoic acid) | 200 | 192 | | ng/L | | 96 | 70 - 130 |
| Perfluoro-3-methoxypropanoic acid (PFMPA) | 200 | 192 | | ng/L | | 96 | 70 - 130 |
| Perfluoro-3,6-dioxaheptanoic acid | 200 | 182 | | ng/L | | 91 | 70 - 130 |

QC Sample Results

Client: Union County Water
Project/Site: PFAS - 533

Job ID: 810-119317-1

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

| Isotope Dilution | LCS LCS | | Limits |
|------------------|-----------|-----------|----------|
| | %Recovery | Qualifier | |
| 13C4 PFBA | 83 | | 50 - 200 |
| 13C5 PFPeA | 83 | | 50 - 200 |
| 13C5 PFHxA | 80 | | 50 - 200 |
| 13C4 PFHpA | 82 | | 50 - 200 |
| 13C8 PFOA | 87 | | 50 - 200 |
| 13C9 PFNA | 85 | | 50 - 200 |
| 13C6 PFDA | 83 | | 50 - 200 |
| 13C7 PFUnA | 82 | | 50 - 200 |
| 13C2 PFDoA | 80 | | 50 - 200 |
| 13C3 HFPO-DA | 78 | | 50 - 200 |
| 13C3 PFBS | 97 | | 50 - 200 |
| 13C8 PFOS | 100 | | 50 - 200 |
| 13C2-4:2-FTS | 104 | | 50 - 200 |
| 13C2-6:2-FTS | 111 | | 50 - 200 |
| 13C2-8:2-FTS | 105 | | 50 - 200 |
| 13C3 PFHxS | 97 | | 50 - 200 |

Lab Sample ID: LLCS 810-114219/2-A

Matrix: Drinking Water

Analysis Batch: 114282

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 114219

| Analyte | Spike Added | LLCS LLCS | | Unit | D | %Rec | %Rec Limits |
|---|-------------|-----------|-----------|------|---|------|-------------|
| | | Result | Qualifier | | | | |
| Perfluorobutanoic acid (PFBA) | 2.00 | 2.07 | | ng/L | | 104 | 50 - 150 |
| Perfluoropentanoic acid (PFPeA) | 2.00 | 1.92 | J | ng/L | | 96 | 50 - 150 |
| Perfluorohexanoic acid (PFHxA) | 2.00 | 1.85 | J | ng/L | | 92 | 50 - 150 |
| Perfluoroheptanoic acid (PFHpA) | 2.00 | 1.93 | J | ng/L | | 97 | 50 - 150 |
| Perfluorooctanoic acid (PFOA) | 2.00 | 1.84 | J | ng/L | | 92 | 50 - 150 |
| Perfluorononanoic acid (PFNA) | 2.00 | 1.92 | J | ng/L | | 96 | 50 - 150 |
| Perfluorodecanoic acid (PFDA) | 2.00 | 1.95 | J | ng/L | | 97 | 50 - 150 |
| Perfluoroundecanoic acid (PFUnA) | 2.00 | 1.88 | J | ng/L | | 94 | 50 - 150 |
| Perfluorododecanoic acid (PFDoA) | 2.00 | 1.89 | J | ng/L | | 95 | 50 - 150 |
| Perfluorobutanesulfonic acid (PFBS) | 1.78 | 1.70 | J | ng/L | | 96 | 50 - 150 |
| Perfluoropentanesulfonic acid (PFPeS) | 1.88 | 1.62 | J | ng/L | | 86 | 50 - 150 |
| Perfluorohexanesulfonic acid (PFHxS) | 1.83 | 1.64 | J | ng/L | | 90 | 50 - 150 |
| Perfluoroheptanesulfonic acid (PFHpS) | 1.91 | 1.66 | J | ng/L | | 87 | 50 - 150 |
| Perfluorooctanesulfonic acid (PFOS) | 1.86 | 1.74 | J | ng/L | | 94 | 50 - 150 |
| Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA) | 1.78 | 1.71 | J | ng/L | | 96 | 50 - 150 |
| 1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS) | 1.88 | 2.12 | | ng/L | | 113 | 50 - 150 |
| 1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS) | 1.90 | 2.21 | | ng/L | | 116 | 50 - 150 |
| 1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS) | 1.92 | 1.87 | J | ng/L | | 98 | 50 - 150 |
| Hexafluoropropylene Oxide Dimer Acid (HFPO-DA) | 2.00 | 1.91 | J | ng/L | | 96 | 50 - 150 |
| 4,8-Dioxa-3H-perfluorononanoic acid (ADONA) | 1.89 | 1.75 | J | ng/L | | 92 | 50 - 150 |

Eurofins Eaton Analytical South Bend

QC Sample Results

Client: Union County Water
 Project/Site: PFAS - 533

Job ID: 810-119317-1

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: LLCS 810-114219/2-A

Matrix: Drinking Water

Analysis Batch: 114282

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 114219

| Analyte | Spike Added | LLCS Result | LLCS Qualifier | Unit | D | %Rec | %Rec Limits |
|--|-------------|-------------|----------------|------|---|------|-------------|
| 9-Chlorohexadecafluoro-3-oxan onane-1-sulfonic acid | 1.87 | 2.11 | | ng/L | | 113 | 50 - 150 |
| 11-Chloroeicosfluoro-3-oxaund ecane-1-sulfonic acid | 1.89 | 1.67 | J | ng/L | | 88 | 50 - 150 |
| Perfluoro(4-methoxybutanoic acid) | 2.00 | 1.88 | J | ng/L | | 94 | 50 - 150 |
| Perfluoro-3-methoxypropanoic acid (PFMPA) | 2.00 | 1.89 | J | ng/L | | 95 | 50 - 150 |
| Perfluoro-3,6-dioxaheptanoic acid | 2.00 | 1.79 | J | ng/L | | 90 | 50 - 150 |

| Isotope Dilution | LLCS LLCS | | Limits |
|------------------|-----------|-----------|----------|
| | %Recovery | Qualifier | |
| 13C4 PFBA | 69 | | 50 - 200 |
| 13C5 PFPeA | 70 | | 50 - 200 |
| 13C5 PFHxA | 71 | | 50 - 200 |
| 13C4 PFHpA | 72 | | 50 - 200 |
| 13C8 PFOA | 76 | | 50 - 200 |
| 13C9 PFNA | 80 | | 50 - 200 |
| 13C6 PFDA | 79 | | 50 - 200 |
| 13C7 PFUnA | 80 | | 50 - 200 |
| 13C2 PFDoA | 82 | | 50 - 200 |
| 13C3 HFPO-DA | 66 | | 50 - 200 |
| 13C3 PFBS | 98 | | 50 - 200 |
| 13C8 PFOS | 98 | | 50 - 200 |
| 13C2-4:2-FTS | 94 | | 50 - 200 |
| 13C2-6:2-FTS | 94 | | 50 - 200 |
| 13C2-8:2-FTS | 98 | | 50 - 200 |
| 13C3 PFHxS | 101 | | 50 - 200 |

QC Association Summary

Client: Union County Water
 Project/Site: PFAS - 533

Job ID: 810-119317-1

LCMS

Prep Batch: 114219

| Lab Sample ID | Client Sample ID | Prep Type | Matrix | Method | Prep Batch |
|---------------------|-----------------------------|-----------|----------------|--------|------------|
| 810-119317-1 | J18 - Rehoboth ARV | Total/NA | Drinking Water | 533 | |
| 810-119317-2 | Y01 - Yadkin Finished Water | Total/NA | Drinking Water | 533 | |
| 810-119317-3 | Y02 - Yadkin Raw Water | Total/NA | Drinking Water | 533 | |
| MBL 810-114219/1-A | Method Blank | Total/NA | Drinking Water | 533 | |
| LCS 810-114219/3-A | Lab Control Sample | Total/NA | Drinking Water | 533 | |
| LLCS 810-114219/2-A | Lab Control Sample | Total/NA | Drinking Water | 533 | |

Analysis Batch: 114282

| Lab Sample ID | Client Sample ID | Prep Type | Matrix | Method | Prep Batch |
|---------------------|-----------------------------|-----------|----------------|--------|------------|
| 810-119317-1 | J18 - Rehoboth ARV | Total/NA | Drinking Water | 533 | 114219 |
| 810-119317-2 | Y01 - Yadkin Finished Water | Total/NA | Drinking Water | 533 | 114219 |
| 810-119317-3 | Y02 - Yadkin Raw Water | Total/NA | Drinking Water | 533 | 114219 |
| MBL 810-114219/1-A | Method Blank | Total/NA | Drinking Water | 533 | 114219 |
| LCS 810-114219/3-A | Lab Control Sample | Total/NA | Drinking Water | 533 | 114219 |
| LLCS 810-114219/2-A | Lab Control Sample | Total/NA | Drinking Water | 533 | 114219 |



Lab Chronicle

Client: Union County Water
Project/Site: PFAS - 533

Job ID: 810-119317-1

Client Sample ID: J18 - Rehoboth ARV

Lab Sample ID: 810-119317-1

Date Collected: 09/09/24 11:10

Matrix: Drinking Water

Date Received: 09/10/24 09:30

| Prep Type | Batch Type | Batch Method | Run | Dilution Factor | Batch Number | Analyst | Lab | Prepared or Analyzed |
|-----------|------------|--------------|-----|-----------------|--------------|---------|-------|----------------------|
| Total/NA | Prep | 533 | | | 114219 | MP | EA SB | 09/11/24 07:53 |
| Total/NA | Analysis | 533 | | 1 | 114282 | MH | EA SB | 09/12/24 03:01 |

Client Sample ID: Y01 - Yadkin Finished Water

Lab Sample ID: 810-119317-2

Date Collected: 09/09/24 11:57

Matrix: Drinking Water

Date Received: 09/10/24 09:30

| Prep Type | Batch Type | Batch Method | Run | Dilution Factor | Batch Number | Analyst | Lab | Prepared or Analyzed |
|-----------|------------|--------------|-----|-----------------|--------------|---------|-------|----------------------|
| Total/NA | Prep | 533 | | | 114219 | MP | EA SB | 09/11/24 07:53 |
| Total/NA | Analysis | 533 | | 1 | 114282 | MH | EA SB | 09/12/24 03:15 |

Client Sample ID: Y02 - Yadkin Raw Water

Lab Sample ID: 810-119317-3

Date Collected: 09/09/24 12:04

Matrix: Drinking Water

Date Received: 09/10/24 09:30

| Prep Type | Batch Type | Batch Method | Run | Dilution Factor | Batch Number | Analyst | Lab | Prepared or Analyzed |
|-----------|------------|--------------|-----|-----------------|--------------|---------|-------|----------------------|
| Total/NA | Prep | 533 | | | 114219 | MP | EA SB | 09/11/24 07:53 |
| Total/NA | Analysis | 533 | | 1 | 114282 | MH | EA SB | 09/12/24 04:22 |

Laboratory References:

EA SB = Eurofins Eaton Analytical South Bend, 110 S Hill Street, South Bend, IN 46617, TEL (574)233-4777

Accreditation/Certification Summary

Client: Union County Water
 Project/Site: PFAS - 533

Job ID: 810-119317-1

Laboratory: Eurofins Eaton Analytical South Bend

Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below.

| Authority | Program | Identification Number | Expiration Date |
|---------------------|---------|-----------------------|-----------------|
| North Carolina (DW) | State | 18700 | 07-31-25 |

The following analytes are included in this report, but the laboratory is not certified by the governing authority. This list may include analytes for which the agency does not offer certification.

| Analysis Method | Prep Method | Matrix | Analyte |
|-----------------|-------------|----------------|---|
| 533 | 533 | Drinking Water | 11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid |
| 533 | 533 | Drinking Water | 1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS) |
| 533 | 533 | Drinking Water | 1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS) |
| 533 | 533 | Drinking Water | 1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS) |
| 533 | 533 | Drinking Water | 4,8-Dioxa-3H-perfluorononanoic acid (ADONA) |
| 533 | 533 | Drinking Water | 9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid |
| 533 | 533 | Drinking Water | Hexafluoropropylene Oxide Dimer Acid (HFPO-DA) |
| 533 | 533 | Drinking Water | Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA) |
| 533 | 533 | Drinking Water | Perfluoro(4-methoxybutanoic acid) |
| 533 | 533 | Drinking Water | Perfluoro-3,6-dioxaheptanoic acid |
| 533 | 533 | Drinking Water | Perfluoro-3-methoxypropanoic acid (PFMPA) |
| 533 | 533 | Drinking Water | Perfluorobutanesulfonic acid (PFBS) |
| 533 | 533 | Drinking Water | Perfluorobutanoic acid (PFBA) |
| 533 | 533 | Drinking Water | Perfluorodecanoic acid (PFDA) |
| 533 | 533 | Drinking Water | Perfluorododecanoic acid (PFDoA) |
| 533 | 533 | Drinking Water | Perfluoroheptanesulfonic acid (PFHpS) |
| 533 | 533 | Drinking Water | Perfluoroheptanoic acid (PFHpA) |
| 533 | 533 | Drinking Water | Perfluorohexanesulfonic acid (PFHxS) |
| 533 | 533 | Drinking Water | Perfluorohexanoic acid (PFHxA) |
| 533 | 533 | Drinking Water | Perfluorononanoic acid (PFNA) |
| 533 | 533 | Drinking Water | Perfluorooctanesulfonic acid (PFOS) |
| 533 | 533 | Drinking Water | Perfluorooctanoic acid (PFOA) |
| 533 | 533 | Drinking Water | Perfluoropentanesulfonic acid (PFPeS) |
| 533 | 533 | Drinking Water | Perfluoropentanoic acid (PFPeA) |
| 533 | 533 | Drinking Water | Perfluoroundecanoic acid (PFUnA) |

Method Summary

Client: Union County Water
Project/Site: PFAS - 533

Job ID: 810-119317-1

| Method | Method Description | Protocol | Laboratory |
|--------|---|----------|------------|
| 533 | Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water | EPA | EA SB |
| 533 | Extraction of Perfluorinated and Polyfluorinated Alkyl Acids | EPA | EA SB |

Protocol References:

EPA = US Environmental Protection Agency

Laboratory References:

EA SB = Eurofins Eaton Analytical South Bend, 110 S Hill Street, South Bend, IN 46617, TEL (574)233-4777



Sample Summary

Client: Union County Water
Project/Site: PFAS - 533

Job ID: 810-119317-1

| Lab Sample ID | Client Sample ID | Matrix | Collected | Received |
|---------------|-----------------------------|----------------|----------------|----------------|
| 810-119317-1 | J18 - Rehoboth ARV | Drinking Water | 09/09/24 11:10 | 09/10/24 09:30 |
| 810-119317-2 | Y01 - Yadkin Finished Water | Drinking Water | 09/09/24 11:57 | 09/10/24 09:30 |
| 810-119317-3 | Y02 - Yadkin Raw Water | Drinking Water | 09/09/24 12:04 | 09/10/24 09:30 |

1

2

3

4

5

6

7

8

9

10

11

12

13

14

15

Chain of Custody Record



| | | | |
|--|---|---|---|
| Client Information Company: Justin Huntley Address: 500 N Main St, Monroe, NC, 28112 Phone: 704-289-3307 (Tel) Email: Justin.Huntley@UnionCountyNC.gov Project Name: PFAS - 533 Site: | | Lab PM: Mattheis, Joe E-Mail: Joe.Mattheis@et.eurofins.com Carrier Tracking No(s): 810-34247-6174.1 State of Origin: | |
| Due Date Requested: TAT Requested (days): Compliance Project: <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No PO #: Purchase Order not required WO #: | | Analysis Requested Barcode: 810-119317 Chain of Custody Preservation Codes: I - NH4 Acetate | |
| Sample Identification JTB - Reobeth ARV Y01 - Yackin Finished Water Y02 - Yackin Raw Water | | Total Number of containers: 3 Special Instructions/Note: Initial Temp: 1.9 Corrected Temp: 2.1 IR Gun # 25177 | |
| Sample Date 9/19/24 9/19/24 9/19/24 | Sample Time 11:10A 11:57A 12:04P | Sample Type (C=Comp, G=grab) G G G | Matrix (W=water, S=solid, O=waste/oil, BT=Tissue, A=air) Drinking Water Drinking Water Drinking Water |
| Possible Hazard Identification <input type="checkbox"/> Non-Hazard <input type="checkbox"/> Flammable <input type="checkbox"/> Skin Irritant <input type="checkbox"/> Poison B <input type="checkbox"/> Unknown <input type="checkbox"/> Radiological | | Sample Disposal (A fee may be assessed if samples are retained longer than 1 month) <input type="checkbox"/> Return To Client <input type="checkbox"/> Disposal By Lab <input type="checkbox"/> Archive For _____ Months | |
| Deliverable Requested: I, II, III, IV, Other (specify) | | Special Instructions/QC Requirements: | |
| Empty Kit Relinquished by: Jordan Helms | | Method of Shipment: | |
| Relinquished by: Date/Time: 9/19/24 2:15P | | Received by: Kameron Williams Date/Time: 9/19/24 0930 | |
| Relinquished by: | | Received by: | |
| Relinquished by: | | Received by: | |
| Custody Seals Intact: <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No | | Cooler Temperature(s) °C and Other Remarks: | |



Login Sample Receipt Checklist

Client: Union County Water

Job Number: 810-119317-1

Login Number: 119317

List Source: Eurofins Eaton Analytical South Bend

List Number: 1

Creator: Williams, Kameron

| Question | Answer | Comment |
|--|--------|---------|
| The cooler's custody seal, if present, is intact. | True | |
| Sample custody seals, if present, are intact. | True | |
| Samples were received on ice. | True | |
| Cooler Temperature is acceptable. | True | |
| Cooler Temperature is recorded. | True | |
| COC is present. | True | |
| COC is filled out in ink and legible. | True | |
| COC is filled out with all pertinent information. | True | |
| There are no discrepancies between the containers received and the COC. | True | |
| Samples are received within Holding Time (excluding tests with immediate HTs) | True | |
| Sample containers have legible labels. | True | |
| Containers are not broken or leaking. | True | |
| Sample collection date/times are provided. | True | |
| There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs | True | |
| Containers requiring zero headspace have no headspace or bubble is <6mm (1/4"). | True | |
| Samples do not require splitting or compositing. | True | |
| Container provided by EEA | True | |

